EU Type Examination Certificate

This is to certify that:

Kunshan Jinfujie Precision Mould Co.,Ltd
Room 78 No.4,Gucheng Middle Road
Yushan Town
Kunshan
Jiangsu
215316
China

Holds Certificate Number: CE 729350

In respect of:

Models JFJ-302 and JFJ-502 Particulate Respirators.
To technical specification Annex II (EHSR) of the PPE Regulation (EU) 2016/425
PPE for use by healthcare professionals as per Commission recommendation 2020/403.

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Drs. Dave Hagenaars, Managing Director

Previous Notified Body: BSI 0086

First Issued: 2020-06-25

Latest Issue: 2020-06-25

Effective Date: 2020-06-25

Expiry Date: 2021-06-25

PRODUCTS RvA C 646

...making excellence a habit.™

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 729350

Product Specification

Product Name:	Particulate Respirators.
Product Type:	Particulate filtering half masks for use by Healthcare professionals.
Models:	**JFJ-302; JFJ-502.**
Classification:	FFP2 NR un-valved.
Technical Specification:	Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.
Product Description:	Both respiratory masks are non-reusable, vertical fold flat type, and secured to the face by a pair of elasticated ear straps, with a plastic buckle to tie the straps at the back of the head. Both models have no exhalation valve. These respirators are FFP2 class. The respirator listed on this certificate is for use by healthcare workers, first responders and other personnel involved in the efforts to contain the COVID-19 virus and avoid its further spread. The product covered by this certificate is not approved for industrial applications and the certificate is only valid as long as EU Commission recommendation sheet 2020/403 remains applicable.
Product Assessments:	BSI's PPE for Healthcare Professionals 2020/403 – RPE Technical Specification.

Customers verify the use of qualifications. Other USES are prohibited. Copy is invalid
仅供验证/禁止用途/否则无效

First Issued: 2020-06-25
Latest Issue: 2020-06-25

Effective Date: 2020-06-25
Expiry Date: 2021-06-25

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 729350

Certificate Administration Details

Technical File Reference: Kunshan Jinfujie Precision Mould Co.,Ltd Technical File.

Certificate Amendment Record:

Issue date	Comments	BSI Review No.
June 2020	First issue.	2797:20:3215335

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type based on Internal Production Control plus supervised product checks at random intervals, Annex VII (Module C2) as referenced on BSI issued Certificate CE 729352.

Customers verify the use of qualifications. Other USES are prohibited. Copy is invalid
仅供验证/禁止商用/复印无效

First Issued: 2020-06-25
Latest Issue: 2020-06-25

Effective Date: 2020-06-25
Expiry Date: 2021-06-25

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.